imegenagro
Food Safety



User Guide

imegen® Pollack ID kit

Pollachius pollachius detection by real time PCR

REF : IMG-223

Instituto de Medicina Genómica SL
Agustín Escardino 9,
Parc Científic de la Universitat de València
46980 Paterna (Valencia, España)
+34 963 212 340 - info@imegen.es

imegen.es

Rev. 1

The information in this guide is subject to change without notice.

Imegen guarantees that its products are free from defects, both in used materials as in its manufacturing process. This warranty is extended to the expiration date, as long as the storage conditions specified in this manual are met. ***Our products are designed for use in research.*** *The user of the product is responsible for validating the usefulness of the protocol proposed by Imegen. These protocols are considered a guide only. Imegen does not offer any other warranty, express or implied, which extend beyond the proper functioning of the components of this set. Imegen sole obligation in respect of the preceding guarantees, will be to replace the product or return the purchase price thereof, as desired by the customer, as long as the existence of a defect in the materials test, or in the manufacture of its products. Imegen will not be responsible for any damage, direct or indirect, resulting in economic losses or damages resulting from the use of this product by the purchaser or user.*

All products sold by the Imegen are subjected to rigorous quality control. The ***imegen® Pollack ID kit*** *has passed all internal validation tests, ensuring the reliability and reproducibility of each assay.*

For any questions about the applications of this product or its protocols, please contact our Technical Department:

Phone number: *963 212 340*
e-Mail: *info@imegen.es*

imegenagro

Index

1. *Product information* 4
   - *Kit description* 4
   - *Content and storage of the kit* 4
   - *Equipment and material required* 5

2. *Real time PCR* 7
   - *PCR reactions preparation* 7
   - *PCR amplification program* 8

3. *Analysis of results* 10

Instituto de Medicina Genómica SL
Agustín Escardino 9,
Parc Científic de la Universitat de València
46980 Paterna (Valencia, España)
+34 963 212 340 - info@imegen.es

imegen.es

ISO 9001
BUREAU VERITAS
Certification
Nº ES063227-1

1. Product information

Kit description

Identification of meat species presence in food samples is an essential step in order to verify the origin and traceability of the used raw materials, as well as a necessary quality control for handling and cleaning processes of production lines.

***Imegen® Pollack ID Kit** allows determining the presence of DNA of Pollachius pollachius in any food.*

Pollachius pollachius DNA detection is done by real time PCR using TaqMan®-MGB probes. One of them, labelled with FAM™ dye, specifically detects one mitochondrial DNA sequence of Pollachius pollachius. The second probe is labelled with VIC® and detects an Internal Positive Control, which is used to rule out inhibitors in the sample and check the correct functioning of the assay.

*The detection limit has been calculated upon standard samples consisting of mixtures of raw Pollachius pollachius meat and other species. **Imegen® Pollack ID Kit** can detect blends containing a percentage over 0.01% of Pollachius pollachius meat. The limit of detection in processed samples varies depending on the composition and food processing.*

*To ensure the representativeness of the results, we recommend the use of a DNA extraction method that allows you to process a large amount of sample (10-20 g). If you do not have a procedure with these features, we recommend the use of **GMO Extraction Kit** (Part No: 4466336).*

Content and storage of the kit

The kit contains needed reagents to perform the real time PCR reactions:

- *Amplification primers for the PCR system.*
- *Two probes labelled with FAM™ and VIC® fluorophores, to detect the*

specific species and the Internal Positive Control (IPC).
- Pollack Positive Control (0,1% of Pollachius pollachius).
- 2X General Master Mix.

The kit contents the necessary reagents to perform 48 reactions with each PCR system:

Reagents	Color	Amount	Storage
Pollack Master-Mix	Red pad	360 µl	-20ºC
Pollack positive control	Orange cap	60 µl	-20ºC
General Master-Mix	White pad	600 µl	4ºC

Table 1. *Kit components and storage temperature of* ***imegen® Pollack ID Kit***

This kit has been proved and is compatible with the following real time PCR platforms: *7500 FAST* and *StepOne Real-Time PCR System* (Thermo Scientific).

Equipment and material required

In the following table the equipment requirements for using ***Imegen® Pollack ID Kit*** are shown:

Equipments	
1	Real-time PCR Thermal Cycler with channels for detection of FAM™ (520 nm) and VIC® (550 nm)
2	Micropipettes (10 µL, 20 µL and 200 µL)
3	Table top centrifuge with adaptors for 96 well PCR plates and/or 0,2 ml tubes
4	Vortex

Materials

1	Optical 96-well reaction plates or 0.2 ml optical tubes
2	Optical adhesive film for 96 well plates or optical caps for 0.2 ml tubes
3	Disposable micropipette filter tips (10 µL, 20 µL and 200 µL)
4	1.5 mL sterile tubes
5	Powder-free latex gloves

2. Real Time PCR

PCR reactions preparation

A PCR master mix should be prepared in order to perform the analysis with ***imegen® Pollack ID kit****:*

The PCR Master Mix contains:

- *Pollack Master-Mix*
- *General Master-Mix (2X)*

The recommended protocol for preparation of amplification reactions is showed below:

1. *Thaw the Pollack Master-Mix, Pollack Positive Control and samples DNA.*
2. *Vortex each reagent and keep cold.*
3. *Add into a 1.5 mL tube, one for each PCR master mix preparation, the following reagents (table 2). To estimate the amount of necessary reagents, we recommend make calculations taking into account the number of samples to be simultaneously analysed, and then considering one more reaction, or increase a 10% the volume of each reagent.*

Reagents	Amount per reaction
Pollack Master-Mix	**7.5 µL**
General Master Mix	**12.5 µL**

Table 2*. Reagents amount per reaction in 7500 FAST or StepOne*

4. *Vortex the PCR Master Mixes tubes and dispense 20 µL into corresponding wells.*

5. *Add 5 µl of each DNA sample at 10 ng/µl, 5 µl of Pollack Positive Control and 5 µl of the Negative Controls* into the appropriate wells.*

   We recommend making each sample analysis in duplicate.

6. *Cover the well plate with optical film or the tubes with optical cover and spin in the centrifuge.*

* *We strongly recommend using an **extraction negative control** for each run of extractions carried out. This control consists in one tube to which no sample is added and which is summited to the same extraction process as the other samples. Likewise, we recommended using a **PCR negative control** for each PCR run; this tube contains no DNA but all PCR reagents.*

PCR amplification program

The following instructions must be followed in order to setup the amplification program:

7500 Fast o StepOne Real-Time PCR system (Thermo Scientific)

- *Experiment type: Quantitation-Standard curve*
- *Ramp rate: standard*
- *Reaction volume: 25 µL*
- *Reference ROX™: include*
- *TaqMan® probes fluorophores:*

Probe	Reporter	Quencher
Pollack	FAM™	MGB
IPC	VIC®	MGB

***Table 3.** Probes information*

- *Optimal program:*

Fields	Step 1 Enzyme activation	Step 2 PCR	
Cycle Number	1 initial cycle	40 cycles	
		Denaturation	Primers binding / Extention
Temperature	95ºC	95ºC	60ºC
Time	10 minutes	15 seconds	1 minute*

***Table 4**. 7500 FAST or StepOne Optimal PCR program*

**Fluorescence detection*

Instituto de Medicina Genómica SL
Agustín Escardino 9,
Parc Científic de la Universitat de València
46980 Paterna (Valencia, España)
+34 963 212 340 - info@imegen.es

imegen.es

ISO 9001
BUREAU VERITAS Certification
Nº ES063227-1

3. Analysis of results

Before analysing the samples results, you should stablish the threshold in 0.4 and, also, you should check that the results obtained in the controls, is as expected:

- ***Positive Control**: The result must always be positive in all amplification reactions, both in the FAM™ channel as VIC®.*
- ***Negative controls**: Amplification must be only detected in the VIC® channel. In this channel an internal positive control (IPC) is detected, which determines the absence of inhibition in the sample.*

IPC

It must be checked that the IPC (VIC®) is positive in all samples, with a Ct similar to the Positive Control. A negative result in the IPC indicates the presence of inhibitors in the sample. It should be noted that IPC result may be negative in samples where a lot of Pollachius pollachius DNA (FAM™) is detected, because the PCR reagents are exhausted before amplification of the IPC begins.

Pollack

Amplification in the FAM™ channel indicates presence of Pollachius Pollachius DNA in the sample.

It is necessary to check if sample Ct is less than the $Ct_{cut\text{-}off}$ in order to determine if one reaction of amplification is positive. Any reaction of amplification with Ct upper than $Ct_{cut\text{-}off}$ may be considered as negative. The $Ct_{cut\text{-}off}$ is equal than the positive control Ct (0.1%) plus 3.32.

$$Ct_{cut\text{-}off} = 3.32 + Ct_{Positive\ Control}$$

Note: Any sample with a Ct equal than $Ct_{cut\text{-}off}$ contains approximately 0.01% of Pollachius pollachius DNA.

In samples where no amplification in the FAM™ channel is seen, we can conclude that no Pollachius pollachius DNA is detected or that their amount in the sample is below than the detection limit.

The following table shows graphically the results that may be obtained from one sample analysis, as well as the interpretation that should be done from the obtained result:

Pollack Master Mix		Interpretation
Pollack	IPC	
-	+	No Pollachius pollachius DNA are detected
+	+	Pollachius pollachius is detected
-	-	PCR inhibitors presence in the sample*
+	-	Sample with big amount of Pollachius pollachius DNA

***Table 5.** Results interpretation*

** If presence of inhibitors in the sample is detected, we recommend checking whether there has been an excess of DNA in the reaction (the recommended maximum is 250 ng). If the amount of DNA is right, we recommend repeating DNA extraction. If the problem persists, please contact our technical department.*

The following table shows graphically the results that may be obtained from the analysis of different assay controls, as well as the interpretation that should be done from the obtained result:

Controls	Pollack Master Mix		Interpretation
	Pollack	IPC	
Positive control	+	+	Expected result
	-	-	PCR Amplification Failure[1]
Extraction negative control	-	+	Expected result
	+	+	Contamination in the DNA extraction procedure[2]
PCR negative control	-	+	Expected result
	+	+	PCR contamination with Pollachius pollachius DNA[3]

Table 6. *Interpretation of possible results*

- *Recommendations:*

[1] ***PCR Amplification Failure:*** *Check amplification program and configuration of fluorescence capture. Amplification failure may be due to a setup technical problem.*

[2] ***Contamination in the DNA extraction procedure:*** *Contamination may be due to some error made in the process of sample handling, reagents contamination, or environmental contamination. Check DNA extraction protocol, wipe the laboratory where DNA extraction process was performed and take care to avoid any contamination during sample homogenization. If necessary, use new aliquots of the reagents used in DNA extraction.*

[3] ***PCR contaminations with Pollachius pollachius DNA:*** *Contamination of PCR reactions may be due to an error made in the process of sample handling, contamination of the reagents or environmental contamination. Thoroughly clean the laboratory where the PCR process was performed, as well as equipment. If necessary, use new aliquots of the reagents used in the PCR. Prepare the PCR reaction containing the Positive Control last to avoid cross contamination.*